

QDec Measurement Summary

Measurement Information




Panel:	Helio_AW40	Meas. Date:	14.06.2016 - 23:38:39
Paneltype:	Heliostat	Operator:	DLR Expert User
Setup:	MeasPosition	Comment:	
Negl. Rim:	0 mm	System:	QDec_2014-101
Orientation:	Mean Slope to Zero	Location:	Jülich

Result Overview

Quality: no specifications Evaluated Surface: 46.8 %

Value	Measurement	Specification	Description
SDx	1.03 mrad		Slope deviation in x-direction (rms)
SDy	0.91 mrad		Slope deviation in y-direction (rms)
ICAP_D1	97.92 %		Intercept on aperture

Result Graphs



Measurement System Information

Config File:	130929_Helio	QDec Version:	3.1.3680 (14.05.2014)
Evaluation Date:	14.06.2016 - 23:52:19		
Last Calibration:	14.08.2014		
PDF Filename:	...\MeasData_160614\Results\Helio\Helio_AW40\Helio_AW40_Rim0_Result_160614235219.pdf		