


QDec Measurement Summary

Measurement Information			
Panel:	Helio_BD72	Meas. Date:	09.07.2024 - 00:40:34
Paneltype:	Heliostat	Operator:	DLR
Setup:	MeasPosition	Comment:	
Negl. Rim:	0 mm	System:	QDec_2014-101
Orientation:	Mean Slope to Zero	Location:	Jylich

Result Overview			
Quality:	no specifications	Evaluated Surface: 97.4 %	
Value	Measurement	Specification	Description
SDx	0.89 mrad		Slope deviation in x-direction (rms)
SDy	0.93 mrad		Slope deviation in y-direction (rms)
ICAP_D1	93.43 %		Intercept on aperture



Measurement System Information			
Config File:	130929_Helio	QDec Version:	1.0.5229 (23.02.2022)
Evaluation Date:	11.07.2024 - 10:06:22		
Last Calibration:	14.08.2014		
PDF Filename:	...\MeasData_240709\Results\Helio\Helio_BD72\Helio_BD72_Rim0_Result_240711100622.pdf		